

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0610M\_cal

System: SMET

Site: Manus

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
05106/08234	31104/58319	WND1	V060901.00	na	na	14/Oct/06	00:30	02:37

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) x 10 <sup>-6</sup> or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
05106/08234	31102/58342	WND1	V060901.00	na	na	

Replacement completed by: Troy Culkan

**NOTES:** Coefficient's for new anemometer have not been entered to loggernet config.

Upper anemometer WND1

Old Part remains on site as spare:

Old part packed for shipment to: Send to Darwin for repair: X

Other old part disposition (explain):

Other comments: